

DECLARATION OF CONFORMITY

Manufacturer MacroArray Diagnostics GmbH (MADx)
Lemböckgasse 59, Top 4
1230 Vienna, Austria

SRN AT-MF-000030541

We declare that this EU declaration of conformity is issued under our sole responsibility and that the below mentioned in vitro diagnostic medical devices meet all the provisions of the Regulation (EU) 2017/746 and the applicable legislation which apply to it.

Products: See Annex 1

Intended Purpose: See Annex 1

Conformity Assessment: In accordance with Reg. (EU) 2017/746 Article 48 (10) and Annex II+III

Notified Body: N/A

Notified Body Number: N/A

Certificates Issued by Notified Body: N/A

Applicable Legislation: The in vitro diagnostic medical devices described above are in conformity with the following legislations:
Regulation (EU) 2017/746



Vienna, Austria
20.05.2026

Dr. Christian Harwanegg,
Chief Executive Officer,
MacroArray Diagnostics GmbH



Macro Array Diagnostics GmbH
Lemböckgasse 59 / Top 4 - 1230 Wien
M. office@macroarraydx.com • P: +43 (0)1 865 25 73
UID: ATU7045105B • FN 448974 g
www.macroarraydx.com

Document Change History

Version	Description	Replaces
1.0	Initial reworking of the document for the IVDR according to the ÄA – 24042020.	-
2.0	German translation, SRN added.	1.0
3.0	Applicable Legislation, CE Marking added, translation moved to another document.	2.0
4.0	No updated in the content of the document, but the document is now based on an officially approved template.	3.0



Annex 1

Product	REF	Risk Classification & Rule	Intended Purpose	Basic UDI-DI	UDI-DI
ConfigXplorer	30-0100-01	Class A, Rule 5a	The ConfigXplorer is an accessory to support the MADx hardware systems. The IVD medical product configures the MADx hardware systems for the ideal camera settings and is used by trained laboratory personnel and medical professionals in a medical laboratory	91201229230JV	9120122923009