



Rango de potencia – Puntas Piezo

		Piezo-Scaler									
		Pyon1		Pyon2		Piezo-Built in		Tigon		Tigon+	
Puntas		Ideal	Máxima	Ideal	Máxima	Ideal	Máxima	Ideal	Máxima	Ideal	Máxima
Detartraje	1U	30	40	30	40	30	40	30	40	30	40
	2U	20	30	20	30	20	30	20	30	20	30
	3U	20	30	20	30	20	30	20	30	20	30
	4U	*	*	30	40	30	40	30	40	30	40
Perio	1P	10	30	10	30	10	30	10	30	10	30
	2Pr / 2PI	*	*	20	30	20	30	20	30	20	30
	3Pr / 3PI	*	*	*	*	*	*	20	30	20	30
Implant	1I	*	*	25	30	25	30	25	30	25	30
Endo	1R	*	*	20	40	20	40	20	40	20	40
	1E	*	*	15	15	15	15	15	15	15	15
	2E	*	*	35	35	35	35	35	35	35	35
	3E	*	*	20	20	20	20	20	20	20	20
Restauración	1R	*	*	35	40	35	40	35	40	35	40
	2R	*	*	35	40	35	40	35	40	35	40
	3Rd / 3Rm	*	*	35	40	35	40	35	40	35	40

* no se libera