




Certificate of Traceability

Level I Laboratory Certification

This certifies that **LKF - Laboratorium fuer Klinische Forschung**, using **Roche cobas c 501 Tina-quant** has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **European Reference Laboratory ESRL#13**

The system evaluated was:

Instrument: cobas c 501	Calibrator Lot: 800807	Application: Tina-quant Hemoglobin A1c III, cobas
Reagent Lot: 855565, 881506	Calibrator Assigned Values: 0.134, 0.345, 0.684, 0.954, 1.303, 1.862 mmol/L HbA1c; 1.436, 7.742 mmol/L Hb	

Date of Certification: October 1, 2025

Certification Expires: October 1, 2026

NGSP Steering Committee Chair

NGSP Network Coordinator

SRL director/ supervisor