




Certificate of Traceability

Level I Laboratory Certification

This certifies that LKF - Laboratorium fuer Klinische Forschung, using Roche cobas c 501 Tina-quant has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the Diabetes Control and Complications Trial Reference method. The comparison was performed with: European Reference Laboratory ESRL#16

The system evaluated was:

Instrument: cobas c 501	Calibrator Lot: 800811	Application: Tina-quant Hemoglobin A1c III, cobas (V-HbA1c / V-Hb)
Reagent Lot: 951365, 954804	Calibrator Assigned Values: 0.128, 0.328, 0.649, 0.905, 1.237, 1.767 mmol/L HbA1c; 1.384, 7.464 mmol/L Hb	*91.5+2.15

Date of Certification: October 1, 2026

Certification Expires: October 1, 2027

Daniel Sauer
NGSP Steering Committee Chair

Christ Rothberg
NGSP Network Coordinator

Carol M. Kelly
SRL director/supervisor